New Sample Company

Calibration History Report

Operator: Bill Smith

Gage ID: **AG-0007** Serial Number: 6546165 Calibrate ID: 115

Description: Micrometer Manufacturer: Mitutoyo

Information: 0-1" OD Micrometer Issued: No

Number of day between calibration: 365

Date Calibrated: 11/8/2026 12:05:42 PM

Department: Quality Assurance

Location: Calibration Room

Conformity Requirement: AS 9201

Work Instruction: WI 7.6.2 Rev: B

Specification: FED SPEC GGG-C-105C

Calibrated In-house: ☑ **Calibrated Outside:** ☐

Calibration Standards:

GB-1000 Gage Blocks Exp. 6/22/2016
Nist#6478923

GB-0001 Gage Balls Exp. 2/27/2016
Nist#984753

Characteristic:	As Found:	As Left:
0.0000	0.0000	0.0000
0.2500	0.2500	0.2500
0.5000	0.5000	0.5000
0.7500	0.7500	0.7500
1.0000	1.0000	1.0000

Temperature 77 °F

Humidity 33%

Status: Active

Internal Note: Replaced worn measuring screw.